

## 10.15 Scope of Accreditation

Issue No: 1/ Issue Date: 02 Dec 2018  
Accreditation Manager: Hamza Khan

### Calibration Laboratory Accreditation No. ACL 0011

is accredited by the GCC Accreditation Center (GAC) in accordance with the recognized International Standard ISO/IEC 17025:2005, "General requirements for the competence of testing and calibration laboratories"

<b>TUV SUD MIDDLE EAST L.L.C Calibration- Laboratory</b>	
<b>Issue No: 1</b>	<b>Issue Date: 02-Dec-2018</b>
Address. <a href="#">JEDDAH, AL MUHAMMADIYA DISTRICT/1, SAFIYA BINT ABDUL MUTTALIB STREET, BESIDE PANDATI, VILLA NO.3582</a>	Contact: <a href="#">Eng. YASSER HASSAN</a> Tel: <a href="#">+966562150277</a> Fax: <a href="#">+966 12 694 6212</a> Email: <a href="mailto:Yasser.hassan@tuvsudme.com">Yasser.hassan@tuvsudme.com</a> Web Address: <a href="http://www.tuv-sud.ae">http://www.tuv-sud.ae</a>

**Locations where calibration activities covered by the above Accreditation Standard are undertaken**

1- [JEDDAH, AL MUHAMMADIYA DISTRICT/1, SAFIYA BINT ABDUL MUTTALIB STREET, BESIDE PANDATI, VILLA NO.3582](#)

## 10.15 Scope of Accreditation

Issue No: 1/ Issue Date: 02 Dec 2018  
Accreditation Manager: Hamza Khan

### For the following scope:

#### Scope:

##### 1. Calibration

1.20 Pressure and vacuum measuring devices

1.80 Calibration of temperature measuring equipment

Scope details are as follows:

#### **Calibration field 1: Pressure**

Measurand / Equipment	Measuring Range	CMC (k=2)	Method (standard/guide + internal procedure)	Permanent lab (P) / Client-site (S) *
Relative pressure (Oil): Manometers with digital or analogue indication and mechanical bourdon tube manometers	0 up to 600 bar	$(2.5 \times 10^{-4} \cdot P + 0.020)$ bar	DKD-R-6-1 (2014) + Internal procedure TUV SUD-TM-01	P

\*: Put only 'P', 'S' or 'P and S'

#### **Calibration field 2: Temperature**

Measurand / Equipment	Measuring Range	CMC (k=2)	Method (standard/guide + internal procedure)	Permanent lab (P) / Client-site (S) *
Temperature measuring chains (Mechanical and digital thermometers)	0 up to 600 °C	1.0 °C	EURAMET cg 8 (2011) + Internal procedure TUV SUD-TM-03	P

\*: Put only 'P', 'S' or 'P and S'

**END**

## 10.15 Scope of Accreditation

Issue No: 1/ Issue Date: 02 Dec 2018  
 Accreditation Manager: Hamza Khan



**TUV SUD MIDDLE EAST L.L.C** is recorded as issuing GAC accredited certificates to organizations in the countries listed below. This list is current at the time of issue of this schedule.

United Arab Emirates	Bahrain	Saudi Arabia	Oman	Qatar	Kuwait	Yemen
√	√	√	√	√	√	√